

Mask/Device Compatibility List

English

This document specifies the masks compatible with ResMed's positive airway pressure devices and the mask setting, if one is required. For full details regarding the correct use of these devices, please refer to the specific User Guides.

Note: Not all products are available in all regions.

Full Face Masks 1

Device	AcuCare F1-4	AirFit [™] F40 ⁴	AirTouch™ F30i ^{1,4}	AirFit™ F30i¹ / AirFit™ F30i Non Magnetic¹	AirFit [™] F30	AirTouch™ F20 / AirTouch™ F20 for Her	AirFit™ F20 /AirFit™ F20 for Her/ AirFit™ F20 Non Magnetic	AirFit™ F10 / AirFit™ F10 for Her
AirCurve [™] 11 VAuto / AirCurve [™] 11 S / AirCurve [™] 11 ST	Χ	Pillows	Full Face	Full Face	Full Face	Full Face	Full Face	Full Face
AirCurve™ 11 ASV PaceWave	Χ	Pillows	Full Face	Full Face	Full Face	Full Face	Full Face	Full Face
AirSense™ 11	N/V	Pillows	Full Face	Full Face	Full Face	Full Face	Full Face	Full Face
AirMini™	Х	Х	X	Х	~	✓	✓	X
AirStart™ 10 / AirSense™ 10	N/V	Pillows	Full Face	Full Face	Full Face	Full Face	Full Face	Full Face
AirCurve [™] 10 VAuto / AirCurve [™] 10 S /AirCurve [™] 10 ST	N/V	Pillows	Full Face	Full Face	Full Face	Full Face	Full Face	Full Face
AirCurve™ 10 CS Pacewave	N/V	Pillows	Full Face	Full Face	Full Face	Full Face	Full Face	Full Face
AirCurve™ 10 CS-A Pacewave	N/V	N/V	N/V	N/V	Full Face	Full Face	Full Face	Full Face
Lumis™	N/V	Pillows ³	Full Face ³	Full Face ³	Full Face ³	Full Face	Full Face	Full Face
S9 AutoSet™ / S9 Elite™	N/V	Pillows	Full Face	Full Face	N/V	Full Face	Full Face	Full Face
S9 Escape™	N/V	Pillows	Full Face	Full Face	N/V	Full Face	Full Face	Full Face
AutoSet CS™ 2	N/V	N/V	N/V	N/V	N/V	N/V	N/V	FULL FACE
AutoSet CS™ (S9)	N/V	Pillows	N/V	N/V	N/V	Full Face	Full Face	Full Face
AutoSet CS™ (S9 PaceWave)	N/V	Pillows	N/V	N/V	N/V	Full Face	Full Face	Full Face
AutoSet CS-A™ (S9 PaceWave)	N/V	N/V	N/V	N/V	N/V	Full Face	Full Face	Full Face
C-Series Tango™	N/V	N/V	N/V	N/V	N/V	N/V	N/V	No setting required. Consult your clinician before changing to a different mask type.
S7™ Lightweight	N/V	N/V	N/V	N/V	N/V	N/V	N/V	No setting required. Consult your clinician before changing to a different mask type.
S7 Elite™ / S7 AutoSet Spirit	N/V	N/V	N/V	N/V	N/V	N/V	N/V	MIR FULL
S7 AutoSet Respond™	N/V	N/V	N/V	N/V	N/V	N/V	N/V	MIRAGE
S8 [™] Auto 25	N/V	N/V	N/V	N/V	N/V	MIR FULL	MIR FULL	MIR FULL
S8 AutoSet Spirit™	N/V	N/V	N/V	N/V	N/V	MIR FULL	MIR FULL	MIR FULL
S8 AutoSet Vantage™	N/V	N/V	N/V	N/V	N/V	MIR FULL	MIR FULL	MIR FULL
S8 Escape [™]	N/V	N/V	N/V	N/V	N/V	N/V	N/V	Requires pressure calibration at 12 cm H2O. Consult your clinician and check the user guide.

Device	AcuCare F1-4	AirFit [™] F40 ⁴	AirTouch [™] F30i ^{1,4}	AirFit [™] F30i ¹ / AirFit [™] F30i Non Magnetic ¹	AirFit [™] F30	AirTouch™ F20 / AirTouch™ F20 for Her	AirFit [™] F20 /AirFit [™] F20 for Her/ AirFit [™] F20 Non Magnetic	AirFit™ F10 / AirFit™ F10 for Her
S8 Escape™ II	N/V	N/V	N/V	N/V	N/V	Requi	ires pressure calibration at 12	cm H2O. Consult your clinician and check the user guide.
S8 Elite™ / S8 Elite™ II	N/V	N/V	N/V	N/V	N/V	MIR FULL	MIR FULL	MIR FULL
S8 AutoSet Spirit™ II	N/V	N/V	N/V	N/V	N/V	MIR FULL	MIR FULL	MIR FULL
S8 Compact [™]	N/V	N/V	N/V	N/V	N/V	N/V	N/V	No setting required. Consult your clinician before changing to a different mask type.
VPAP™ III / VPAP™ III ST	N/V	N/V	N/V	N/V	N/V	N/V	N/V	MIR FULL
VPAP™ III ST-A	N/V	N/V	N/V	N/V	N/V	N/V	N/V	MIR FULL
VPAP™ IV / VPAP™ IV ST	N/V	N/V	N/V	N/V	N/V	MIR FULL	MIR FULL	MIR FULL
VPAP™ Adapt SV	N/V	N/V	N/V	N/V	N/V	N/V	N/V	FULL FACE
VPAP™ Adapt (S9)	N/V	Pillows	Full Face	Full Face	N/V	Full Face	Full Face	Full Face
VPAP™ Adapt (S9 PaceWave)	N/V	Pillows	N/V	Full Face	N/V	Full Face	Full Face	Full Face
VPAP™ Malibu	N/V	N/V	N/V	N/V	N/V	N/V	N/V	FULL FACE
VPAP™ Auto 25	N/V	N/V	N/V	N/V	N/V	MIR FULL	MIR FULL	MIR FULL
Auto 25 (S9)	N/V	Pillows	N/V	Full Face	N/V	Full Face	Full Face	Full Face
VPAP™S / VPAP™ST	N/V	N/V	N/V	N/V	N/V	MIR FULL	MIR FULL	MIR FULL
VPAP™ S (S9) / VPAP™ ST (S9)	N/V	Pillows	Full Face	Full Face	N/V	Full Face	Full Face	Full Face
VPAP™ ST-A (S9)	N/V	N/V	N/V	N/V	N/V	Full Face	Full Face	Full Face
VPAP™ ST (S9 iVAPS) / VPAP™ ST-A (S9 iVAPS)	N/V	N/V	N/V	N/V	N/V	Full Face	Full Face	Full Face
S9 VPAP™ Tx	N/V	Pillows ²	Full Face ²	Full Face ²	N/V	Full Face	Full Face	Full Face
VPAP™ Tx	N/V	N/V	N/V	N/V	N/V	N/V	N/V	FULL FACE
Stellar™	N/V	N/V	N/V	N/V	N/V	FULL FACE	FULL FACE	FULL FACE
VS Ultra™ / VS Integra™	~	N/V	N/V	N/V	N/V	N/V	N/V	✓
VS III™	N/V	N/V	N/V	N/V	N/V	N/V	N/V	✓
Astral™	FULL FACE	N/V	N/V	N/V	N/V	FULL FACE	FULL FACE	N/V

¹ SmartStop may not operate as described when using this mask with some CPAP or bi-level devices.

Full Face Masks 2

Device	Mirage™	Mirage Series 2	Ultra Mirage™	Mirage Quattro™	Quattro™ FX / Quattro™ FX for Her Quattro™ Air / Quattro™ Air for Her	Mirage Liberty [™]	Hospital Full Face	
AirCurve™ 11 VAuto / AirCurve™ 11 S / AirCurve™ 11 ST	N/V	N/V	Full Face	Full Face	Full Face	Full Face	Full Face	
AirCurve™ 11 ASV PaceWave	N/V	N/V	Full Face	Full Face	Full Face	Full Face	Full Face	
AirSense™ 11	N/V	N/V	Full Face	Full Face	Full Face	Full Face	Full Face	
AirMini™	Χ	Х	X	X	X	X	X	

² Excluding PAC and iVAPS modes. Only compatible for pressure range 4 - 25 cm H2O.

³ This combination of mask and device has been determined safe for use based on the evaluation of accuracy of monitoring of pressure, volume, minute ventilation and respiratory rate, and its alarms if applicable. Small variations in performance might be observed, particularly at higher pressure or flow. As per expected clinical practice, the effectiveness of ventilation and alarms should be verified by the clinician.

⁴ SmartStart may not operate as described when using this mask with some CPAP or bi-level devices.

N/V = The mask has not been validated for this device.

 $[\]checkmark$ = Compatible X = Not Compatible

Device	Mirage™	Mirage Series 2	Ultra Mirage [™]	Mirage Quattro™	Quattro™ FX / Quattro™ FX for Her Quattro™ Air / Quattro™ Air for Her	Mirage Liberty [™]	Hospital Full Face
AirStart™ 10 / AirSense™ 10	N/V	N/V	Full Face	Full Face	Full Face	Full Face	Full Face
AirCurve [™] 10 VAuto / AirCurve [™] 10 S / AirCurve [™] 10 ST	N/V	N/V	Full Face	Full Face	Full Face	Full Face	Full Face
AirCurve™ 10 CS Pacewave	N/V	N/V	Full Face	Full Face	Full Face	Full Face	Full Face
AirCurve™ 10 CS-A Pacewave	N/V	N/V	Full Face	Full Face	Full Face	Full Face	Full Face
Lumis™	N/V	N/V	Full Face	Full Face	Full Face	Full Face	Full Face
S9 AutoSet [™] / S9 Elite [™]	N/V	N/V	Full Face	Full Face	Full Face	Full Face	Full Face
S9 Escape [™]	N/V	N/V	Full Face	Full Face	Full Face	Full Face	Full Face
AutoSet CS™ 2	FULL FACE	FULL FACE	FULL FACE	FULL FACE	FULL FACE	FULL FACE ¹	FULL FACE
AutoSet CS™ (S9)	N/V	N/V	Full Face	Full Face	Full Face	Full Face	Full Face
AutoSet CS™ (S9 PaceWave)	N/V	N/V	Full Face	Full Face	Full Face	Full Face	Full Face
AutoSet CS-A™ (S9 PaceWave)	N/V	N/V	Full Face	Full Face	Full Face	Full Face	Full Face
C-Series Tango™				No setting required. Consult yo	our clinician before changing to a different mask ty	pe.	
S7™ Lightweight				No setting required. Consult yo	our clinician before changing to a different mask ty	pe.	
S7 Elite™ / S7 AutoSet Spirit	MIR FULL	MIR FULL	MIR FULL	MIR FULL	MIR FULL	MIR FULL	MIR FULL
S7 AutoSet Respond™	STANDARD	ULTRA	MIRAGE	MIRAGE	MIRAGE	MIR FULL	MIR FULL
S8 [™] Auto 25	MIR FULL	MIR FULL	MIR FULL	MIR FULL	MIR FULL	MIR FULL	MIR FULL
S8 AutoSet Spirit™	MIR FULL	MIR FULL	MIR FULL	MIR FULL	MIR FULL	MIR FULL	MIR FULL
S8 AutoSet Vantage™	MIR FULL	MIR FULL	MIR FULL	MIR FULL	MIR FULL	MIR FULL	MIR FULL
S8 Escape [™]			Rec	uires pressure calibration at 12	cm H2O. Consult your clinician and check the user	guide.	
S8 Escape™ II			Rec	uires pressure calibration at 12	cm H2O. Consult your clinician and check the user	guide.	
S8 Elite [™] / S8 Elite [™] II	MIR FULL	MIR FULL	MIR FULL	MIR FULL	MIR FULL	MIR FULL	MIR FULL
S8 AutoSet Spirit™ II	MIR FULL	MIR FULL	MIR FULL	MIR FULL	MIR FULL	MIR FULL	MIR FULL
S8 Compact™				No setting required. Consult yo	our clinician before changing to a different mask ty	pe.	
VPAP™ III / VPAP™ III ST	MIR FULL	MIR FULL	MIR FULL	MIR FULL	MIR FULL	MIR FULL	MIR FULL
VPAP™ III ST-A	MIR FULL	MIR FULL	MIR FULL	MIR FULL	MIR FULL	MIR FULL	MIR FULL
VPAP™ IV / VPAP™ IV ST	MIR FULL	MIR FULL	MIR FULL	MIR FULL	MIR FULL	MIR FULL	MIR FULL
VPAP™ Adapt SV	FULL FACE	FULL FACE	FULL FACE	FULL FACE	FULL FACE	FULL FACE ¹	FULL FACE
VPAP™ Adapt (S9)	N/V	N/V	Full Face	Full Face	Full Face	Full Face	Full Face
VPAP™ Adapt (S9 PaceWave)	N/V	N/V	Full Face	Full Face	Full Face	Full Face	Full Face
VPAP™ Malibu	FULL FACE	FULL FACE	FULL FACE	FULL FACE	FULL FACE	FULL FACE	FULL FACE
VPAP™ Auto 25	MIR FULL	MIR FULL	MIR FULL	MIR FULL	MIR FULL	MIR FULL	MIR FULL
Auto 25 (S9)	N/V	N/V	Full Face	Full Face	Full Face	Full Face	Full Face
VPAP™ S / VPAP™ ST	MIR FULL	MIR FULL	MIR FULL	MIR FULL	MIR FULL	MIR FULL	MIR FULL
VPAP™ S (S9) / VPAP™ ST (S9)	N/V	N/V	Full Face	Full Face	Full Face	Full Face	Full Face
VPAP™ ST-A (S9)	N/V	N/V	Full Face	Full Face	Full Face	Full Face	Full Face
VPAP™ ST (S9 iVAPS) / VPAP™ ST-A (S9 iVAPS)	N/V	N/V	Full Face	Full Face	Full Face	Full Face	Full Face
S9 VPAP™ Tx	N/V	N/V	Full Face	Full Face	Full Face	Full Face	Full Face
VPAP™ Tx	N/V	N/V	FULL FACE	FULL FACE	FULL FACE	FULL FACE	FULL FACE
Stellar™	N/V	N/V	FULL FACE	FULL FACE	FULL FACE	FULL FACE	FULL FACE

Device	Mirage™	Mirage [™] Mirage Series 2		Mirage Quattro™	Quattro™ FX / Quattro™ FX for Her Quattro™ Air / Quattro™ Air for Her	Mirage Liberty [™]	Hospital Full Face
VS Ultra [™] / VS Integra [™]	N/V	N/V	✓	✓	N/V	✓	✓
VS III™	N/V	N/V	✓	✓	N/V	✓	✓
Astral™	N/V	N/V	FULL FACE	FULL FACE	FULL FACE	FULL FACE	FULL FACE

Nasal Masks 1

Device	AirTouch [™] N30i'	AirFit [™] N30 for AirMini	AirFit [™] N30i ¹ / AirFit [™] N30	AirFit™ N20 Classic	AirFit™ N20 / AirFit™ N20 for Her	AirTouch™ N20 / AirTouch™ N20 for Her	AirFit™ N10 / AirFit™ N10 for Her	Swift™ FX Nano / Swift™ FX Nano for Her	Mirage™ FX / Mirage FX™ for Her	Mirage Activa™ LT	Mirage™ SoftGel	Hospital Nasal
AirCurve [™] 11 VAuto / AirCurve [™] 11 S / AirCurve [™] 11 ST	Pillows	Х	Pillows	Pillows	Pillows	Pillows	Pillows	Pillows	Nasal	Nasal	Х	Nasal
AirCurve™ 11 ASV PaceWave	Pillows	Х	Pillows	Pillows	Pillows	Pillows	Pillows	Pillows	Nasal	Nasal	Χ	Nasal
AirSense™ 11	Pillows	Χ	Pillows	Pillows	Pillows	Pillows	Pillows	Pillows	Nasal	Nasal	N/V	Nasal
AirMini™	Χ	✓	Χ	~	~	✓	Χ	Χ	Χ	Χ	Χ	Χ
AirStart™ 10 / AirSense™ 10	Pillows	Χ	Pillows	Pillows	Pillows	Pillows	Pillows	Pillows	Nasal	Nasal	Nasal	Nasal
AirCurve™ 10 VAuto / AirCurve™ 10 S / AirCurve™ 10 ST	Pillows	Χ	Pillows	Pillows	Pillows	Pillows	Pillows	Pillows	Nasal	Nasal	Nasal	Nasal
AirCurve™ 10 CS Pacewave	Pillows	Х	Pillows	Pillows	Pillows	Pillows	Pillows	Pillows	Nasal	Nasal	Nasal	Nasal
AirCurve™ 10 CS-A Pacewave	N/V	Χ	N/V	Pillows	Pillows	Pillows	Pillows	Pillows	Nasal	Nasal	Nasal	Nasal
Lumis™	Pillows ³	Χ	Pillows ³	Pillows	Pillows	Pillows	Pillows	Pillows	Nasal	Nasal	Nasal	Nasal
S9 AutoSet [™] / S9 Elite [™]	Pillows	Χ	Pillows	Pillows	Pillows	Pillows	Pillows	Pillows	Nasal	Nasal	Nasal	Nasal
S9 Escape [™]	Pillows	Χ	Pillows	Pillows	Pillows	Pillows	Pillows	Pillows	Nasal	Nasal	Nasal	Nasal
AutoSet CS™ 2	N/V	Χ	N/V	N/V	N/V	N/V	N/V	N/V	VISTA ¹	ACTIVA	ACTIVA	VISTA
AutoSet CS™ (S9)	N/V	Χ	Pillows	Pillows	Pillows	Pillows	Pillows	Pillows	Nasal	Nasal	Nasal	Nasal
AutoSet CS™ (S9 PaceWave)	N/V	Χ	Pillows	Pillows	Pillows	Pillows	Pillows	Pillows	Nasal	Nasal	Nasal	Nasal
AutoSet CS-A [™] (S9 PaceWave)	N/V	Х	N/V	Pillows	Pillows	Pillows	Pillows	Pillows	Nasal	Nasal	Nasal	Nasal
C-Series Tango™	N/V	Χ	N/V	N/V	N/V	N/V	N/V	N/V	No setting required. C	onsult your cliniciar	before changing to	o a different mask type.
S7™ Lightweight	N/V	Χ	N/V	N/V	N/V	N/V	N/V	N/V	No setting required. C	onsult your cliniciar	before changing to	o a different mask type.
S7 Elite™ / S7 AutoSet Spirit	N/V	Χ	N/V	N/V	N/V	N/V	N/V	N/V	STANDARD	STANDARD	STANDARD	STANDARD
S7 AutoSet Respond™	N/V	Χ	N/V	N/V	N/V	N/V	N/V	N/V	STANDARD	STANDARD	STANDARD	STANDARD
S8™ Auto 25	N/V	Х	N/V	N/V	SWIFT	SWIFT	SWIFT	SWIFT	STANDARD	ACTIVA	ACTIVA	STANDARD
S8 AutoSet Spirit™	N/V	Χ	N/V	N/V	SWIFT	SWIFT	SWIFT	SWIFT	STANDARD	ACTIVA	ACTIVA	STANDARD
S8 AutoSet Vantage™	N/V	Х	N/V	N/V	SWIFT	SWIFT	SWIFT	SWIFT	STANDARD	ACTIVA	ACTIVA	STANDARD
S8 Escape™	N/V	Χ	N/V	N/V	N/V	N/V	Requir	es pressure calibratior	n at 12 cm H2O. Cons	ult your clinician	and check the u	ser guide.
S8 Escape™ II	N/V	Χ	N/V	N/V		Requires press	ure calibration at	12 cm H2O. Consult yo	ur clinician and chec	k the user guide.		
S8 Elite [™] / S8 Elite [™] II	N/V	Χ	N/V	N/V	SWIFT	SWIFT	SWIFT	SWIFT	STANDARD	ACTIVA	ACTIVA	STANDARD
S8 AutoSet Spirit™ II	N/V	Χ	N/V	N/V	SWIFT	SWIFT	SWIFT	SWIFT	STANDARD	ACTIVA	ACTIVA	STANDARD
S8 Compact™	N/V	Χ	N/V	N/V	N/V	N/V	No setting required. Consult your clinician before changing to a different mask type.					
VPAP™ III / VPAP™ III ST	N/V	Х	N/V	N/V	N/V	N/V	MIRAGE	MIRAGE	STANDARD	STANDARD	STANDARD	N/V
VPAP™ III ST-A	N/V	Χ	N/V	N/V	N/V	N/V	MIRAGE	MIRAGE	STANDARD	STANDARD	STANDARD	N/V
VPAP™ IV / VPAP™ IV ST	N/V	Χ	N/V	N/V	SWIFT	SWIFT	SWIFT	SWIFT	STANDARD	ACTIVA	ACTIVA	STANDARD
VPAP™ Adapt SV	N/V	Χ	N/V	N/V	N/V	N/V	N/V	N/V	VISTA ¹	ACTIVA	ACTIVA	VISTA

Device	AirTouch [™] N30i¹	AirFit [™] N30 for AirMini	AirFit [™] N30i ¹ / AirFit [™] N30	AirFit [™] N20 Classic	AirFit™ N20 / AirFit™ N20 for Her	AirTouch™ N20 / AirTouch™ N20 for Her	AirFit [™] N10 / AirFit [™] N10 for Her	Swift™ FX Nano / Swift™ FX Nano for Her	Mirage [™] FX / Mirage FX [™] for Her	Mirage Activa™ LT	Mirage [™] SoftGel	Hospital Nasal
VPAP™ Adapt (S9)	N/V	Χ	Pillows	Pillows	Pillows	Pillows	Pillows	Pillows	Nasal	Nasal	Nasal	Nasal
VPAP™ Adapt (S9 PaceWave)	N/V	Χ	Pillows	Pillows	Pillows	Pillows	Pillows	Pillows	Nasal	Nasal	Nasal	Nasal
VPAP™ Malibu	N/V	Χ	N/V	N/V	N/V	N/V	SWIFT	SWIFT	VISTA ¹	N/V	N/V	MERIDIAN
VPAP™ Auto 25	N/V	Х	N/V	N/V	SWIFT	SWIFT	SWIFT	SWIFT	STANDARD	ACTIVA	ACTIVA	STANDARD
Auto 25 (S9)	N/V	Χ	Pillows	Pillows	Pillows	Pillows	Pillows	Pillows	Nasal	Nasal	Nasal	Nasal
VPAP™ S / VPAP™ ST	N/V	Χ	N/V	N/V	SWIFT	SWIFT	SWIFT	SWIFT	STANDARD	ACTIVA	ACTIVA	STANDARD
VPAP™ S (S9) / VPAP™ ST (S9)	Pillows	Χ	Pillows	Pillows	Pillows	Pillows	Pillows	Pillows	Nasal	Nasal	Nasal	Nasal
VPAP™ ST-A (S9)	N/V	Χ	N/V	Pillows	Pillows	Pillows	Pillows	Pillows	Nasal	Nasal	Nasal	Nasal
VPAP™ ST (S9 iVAPS) / VPAP™ ST-A (S9 iVAPS)	N/V	Х	N/V	Pillows	Pillows	Pillows	Pillows	Pillows	Nasal	Nasal	Nasal	Nasal
S9 VPAP™ Tx	Pillows ²	Χ	Pillows ²	N/V	Pillows	N/V	Pillows	Pillows	Nasal	Nasal	Nasal	Nasal
VPAP™ Tx	N/V	Χ	N/V	N/V	N/V	N/V	PILLOWS	PILLOWS	NASAL	NASAL	NASAL	NASAL
Stellar™	N/V	Χ	N/V	PILLOWS	PILLOWS	PILLOWS	PILLOWS ¹	PILLOWS ¹	NASAL	NASAL	NASAL	NASAL
VS Ultra [™] / VS Integra [™]	N/V	Χ	N/V	N/V	N/V	N/V	N/V	N/V	N/V	N/V	N/V	N/V
VS III™	N/V	Χ	N/V	N/V	N/V	N/V	N/V	N/V	N/V	✓	N/V	~
Astral™	N/V	Χ	N/V	Pillows	Pillows	Pillows	N/V	N/V	Nasal	N/V	Nasal	Nasal

¹ SmartStop may not operate as described when using this mask with some CPAP or bi-level devices.

Nasal Masks 2

Device	Pixi™	Mirage Kidsta™	Mirage Micro™	Mirage Vista™	Ultra Mirage II	Mirage	Mirage Activa™	Modular	Papillon™	Meridian
AirCurve™ 11 VAuto / AirCurve™ 11 S / AirCurve™ 11 ST	Pediatric	Х	X	Х	X	N/V	Х	N/V	N/V	Х
AirCurve™ 11 ASV PaceWave	N/V	Χ	Χ	Χ	X	N/V	Χ	N/V	N/V	X
AirSense™ 11	N/V	N/V	N/V	N/V	N/V	N/V	N/V	N/V	N/V	N/V
AirMini™	Χ	Χ	X	X	X	Х	X	Х	Х	X
AirStart™ 10 / AirSense™ 10	N/V	Nasal	Nasal	Nasal	Nasal	N/V	Nasal	N/V	N/V	Nasal
AirCurve [™] 10 VAuto / AirCurve [™] 10 S / AirCurve [™] 10 ST	N/V	Nasal	Nasal	Nasal	Nasal	N/V	Nasal	N/V	N/V	Nasal
AirCurve™ 10 CS Pacewave	N/V	Nasal	Nasal	Nasal	Nasal	N/V	Nasal	N/V	N/V	Nasal
AirCurve™ 10 CS-A Pacewave	N/V	Nasal	Nasal	Nasal	Nasal	N/V	Nasal	N/V	N/V	Nasal
Lumis™	N/V	Nasal	Nasal	Nasal	Nasal	N/V	Nasal	N/V	N/V	Nasal
S9 AutoSet [™] / S9 Elite [™]	N/V	N/V	Nasal	Nasal	Nasal Ultra	N/V	Nasal	Nasal	Nasal	Nasal
S9 Escape [™]	N/V	N/V	Nasal	Nasal	Nasal Ultra	N/V	Nasal	Nasal	Nasal	Nasal
AutoSet CS™ 2	N/V	N/V	VISTA ¹	VISTA	ULTRA	N/V	ACTIVA	N/V	PAPILLON	VISTA
AutoSet CS™ (S9)	N/V	N/V	Nasal	Nasal	Nasal Ultra	N/V	Nasal	Nasal	Nasal	Nasal

² Excluding PAC and iVAPS modes. Only compatible for pressure range 4 - 25 cm H2O.

³ This combination of mask and device has been determined safe for use based on the evaluation of accuracy of monitoring of pressure, volume, minute ventilation and respiratory rate, and its alarms if applicable. Small variations in performance might be observed, particularly at higher pressure or flow. As per expected clinical practice, the effectiveness of ventilation and alarms should be verified by the clinician.

N/V = The mask has not been validated for this device.

 $[\]checkmark$ = Compatible X = Not Compatible

Device	$Pixi^{TM}$	Mirage Kidsta [™]	Mirage Micro [™]	Mirage Vista™	Ultra Mirage II	Mirage	Mirage Activa [™]	Modular	Papillon™	Meridian
AutoSet CS™ (S9 PaceWave) / AutoSet CS-A™ (S9 PaceWave)	N/V	N/V	Nasal	Nasal	Nasal Ultra	N/V	Nasal	Nasal	Nasal	Nasal
C-Series Tango™	N/V	N/V	N/V		No setting required.	Consult your clinici	an before changing to a dif	ferent mask type.		N/V
S7™ Lightweight	N/V	N/V	N/V		No setting required.	Consult your clinici	an before changing to a dif	ferent mask type.		N/V
S7 Elite™ / S7 AutoSet Spirit	N/V	N/V	STANDARD	STANDARD	ULTRA	MIRAGE	STANDARD	STANDARD	STANDARD	N/V
S7 AutoSet Respond™	N/V	N/V	STANDARD	STANDARD	ULTRA	MIRAGE	STANDARD	STANDARD	STANDARD	N/V
S8™ Auto 25	N/V	N/V	STANDARD	STANDARD	ULTRA	MIRAGE	ACTIVA	N/V	N/V	STANDARD
S8 AutoSet Spirit™	N/V	N/V	STANDARD	STANDARD	ULTRA	MIRAGE	ACTIVA	STANDARD	STANDARD	STANDARD
S8 AutoSet Vantage™	N/V	N/V	STANDARD	STANDARD	ULTRA	MIRAGE	ACTIVA	STANDARD	STANDARD	STANDARD
S8 Escape [™]	N/V	N/V			Requires pressure calibr	ation at 12 cm H20	. Consult your clinician and	check the user guide.		
S8 Escape™ II	N/V	N/V			Requires pressure calibr	ation at 12 cm H2O	. Consult your clinician and	check the user guide.		
S8 Elite™ / S8 Elite™ II	N/V	N/V	STANDARD	STANDARD	ULTRA	MIRAGE	ACTIVA	STANDARD	STANDARD	STANDARD
S8 AutoSet Spirit™ II	N/V	N/V	STANDARD	STANDARD	ULTRA	MIRAGE	ACTIVA	STANDARD	STANDARD	STANDARD
S8 Compact [™]	N/V	N/V			No setting required.	Consult your clinici	an before changing to a dif	ferent mask type.		
VPAP™ III / VPAP™ III ST	N/V	N/V	STANDARD	STANDARD	ULTRA	MIRAGE	STANDARD	STANDARD	N/V	N/V
VPAP™ III ST-A	N/V	STANDARD	STANDARD	STANDARD	ULTRA	MIRAGE	STANDARD	STANDARD	N/V	N/V
VPAP™ IV / VPAP™ IV ST	N/V	N/V	STANDARD	STANDARD	ULTRA	MIRAGE	ACTIVA	N/V	N/V	STANDARD
VPAP™ Adapt SV	N/V	N/V	VISTA ¹	VISTA	ULTRA	N/V	ACTIVA	N/V	N/V	VISTA
VPAP™ Adapt (S9)	N/V	N/V	Nasal	Nasal	Nasal Ultra	N/V	Nasal	Nasal	Nasal	Nasal
VPAP™ Adapt (S9 PaceWave)	N/V	N/V	Nasal	Nasal	Nasal Ultra	N/V	Nasal	Nasal	Nasal	Nasal
VPAP™ Malibu	N/V	N/V	MERIDIAN	VISTA	ULTRA	N/V	N/V	N/V	N/V	MERIDIAN
VPAP™ Auto 25	N/V	N/V	STANDARD	STANDARD	ULTRA	MIRAGE	ACTIVA	N/V	N/V	STANDARD
Auto 25 (S9)	N/V	N/V	Nasal	Nasal	Nasal Ultra	N/V	Nasal	Nasal	Nasal	STANDARD
VPAP™ S / VPAP™ ST	N/V	N/V	STANDARD	STANDARD	ULTRA	MIRAGE	ACTIVA	N/V	N/V	STANDARD
VPAP™ S (S9) / VPAP™ ST (S9) / VPAP™ ST-A (S9)	Pediatric	N/V	Nasal	Nasal	Nasal Ultra	N/V	Nasal	Nasal	Nasal	Nasal
VPAP™ ST (S9 iVAPS) / VPAP™ ST-A (S9 iVAPS)	Pediatric	N/V	Nasal	Nasal	Nasal Ultra	N/V	Nasal	Nasal	Nasal	Nasal
S9 VPAP™ Tx	Pediatric	N/V	Nasal	Nasal	Nasal Ultra	N/V	Nasal	N/V	N/V	Nasal
VPAP™ Tx	N/V	NASAL	NASAL	NASAL	NASAL-ULTRA	N/V	NASAL	N/V	N/V	NASAL
Stellar™	Pediatric	NASAL	NASAL	NASAL	ULTRA	N/V	NASAL	N/V	N/V	N/V
VS Ultra [™] / VS Integra [™]	N/V	N/V	N/V	N/V	N/V	N/V	N/V	N/V	N/V	N/V
VS III™	N/V	✓	~	~	✓	N/V	~	N/V	N/V	N/V
Astral™	Pediatric	Nasal	Nasal	Nasal	Nasal Ultra	N/V	N/V	N/V	N/V	N/V

N/V = The mask has not been validated for this device.

 $[\]checkmark$ = Compatible X = Not Compatible

Nasal Pillows

Device	AirFit [™] P30i ¹	AirFit™ P10 For AirMini™	AirFit™ P10 / AirFit™ P10 for Her	Mirage Swift™	Mirage Swift™ II	Swift™ FX / Swift™ FX for Her / Swift™ FX Bella / Swift™ FX Bella Gray	Swift [™] LT	Swift™ LT for Her / Swift™ LT-F
AirCurve™ 11 VAuto / AirCurve™ 11 S / AirCurve™ 11 ST	Pillows	X	Pillows ²	N/V	N/V	Pillows	Pillows	Pillows
AirCurve™ 11 ASV PaceWave	Pillows	Χ	Pillows ²	N/V	N/V	Pillows	Pillows	Pillows
AirSense [™] 11	Pillows	Χ	Pillows	N/V	Pillows	Pillows	Pillows	Pillows
AirMini™	Χ	✓	Χ	Χ	Χ	X	Х	Х
AirStart [™] 10 / AirSense [™] 10	Pillows	Χ	Pillows ²	N/V	Pillows	Pillows	Pillows	Pillows
AirCurve [™] 10 VAuto / AirCurve [™] 10 S / AirCurve [™] 10 ST	Pillows	X	Pillows ²	N/V	Pillows	Pillows	Pillows	Pillows
AirCurve™ 10 CS Pacewave	Pillows	Х	Pillows ²	N/V	Pillows	Pillows	Pillows	Pillows
AirCurve™ 10 CS-A Pacewave	N/V	Х	Pillows ²	N/V	Pillows	Pillows	Pillows	Pillows
Lumis™	Pillows ⁴	Х	Pillows ²	N/V	Pillows	Pillows	Pillows	Pillows
S9 AutoSet [™] / S9 Elite [™]	Pillows	Χ	Pillows	Pillows	Pillows	Pillows	Pillows	Pillows
S9 Escape [™]	Pillows	Х	Pillows	Pillows	Pillows	Pillows	Pillows	Pillows
AutoSet CS™ 2	N/V	Χ	N/V	N/V	N/V	N/V	N/V	N/V
AutoSet CS™ (S9)	Pillows	Χ	Pillows ²	Pillows	Pillows	Pillows	Pillows	Pillows
AutoSet CS™ (S9 PaceWave)	Pillows	Χ	Pillows ²	Pillows	Pillows	Pillows	Pillows	Pillows
AutoSet CS-A™ (S9 PaceWave)	N/V	Χ	Pillows ²	Pillows	Pillows	Pillows	Pillows	Pillows
C-Series Tango™	N/V	Χ			No setting required. Consu	It your clinician before changing to a different mask	type.	
S7™ Lightweight	N/V	Χ			No setting required. Consu	It your clinician before changing to a different mask	type.	
S7 Elite™ / S7 AutoSet Spirit	N/V	Χ	MIRAGE	MIRAGE	MIRAGE	MIRAGE	MIRAGE	MIRAGE
S7 AutoSet Respond™	N/V	Χ	MIRAGE	MIRAGE	MIRAGE	MIRAGE	MIRAGE	MIRAGE
S8 [™] Auto 25	N/V	Χ	SWIFT	SWIFT	SWIFT	SWIFT	SWIFT	SWIFT
S8 AutoSet Spirit™	N/V	Χ	SWIFT	SWIFT	SWIFT	SWIFT	SWIFT	SWIFT
S8 AutoSet Vantage™	N/V	Χ	SWIFT	SWIFT	SWIFT	SWIFT	SWIFT	SWIFT
S8 Escape [™]	N/V	Χ		Requ	uires pressure calibration a	t 12 cm H2O. Consult your clinician and check the us	er guide.	
S8 Escape™ II	N/V	Χ		Requ	uires pressure calibration a	t 12 cm H2O. Consult your clinician and check the us	er guide.	
S8 Elite™ / S8 Elite™ II	N/V	Χ	SWIFT	SWIFT	SWIFT	SWIFT	SWIFT	SWIFT
S8 AutoSet Spirit™ II	N/V	X	SWIFT	SWIFT	SWIFT	SWIFT	SWIFT	SWIFT
S8 Compact™	N/V	X			No settina required. Consu	It your clinician before changing to a different mask		
VPAP™ III / VPAP™ III ST	N/V	X	MIRAGE ²	MIRAGE	MIRAGE	MIRAGE	MIRAGE	MIRAGE
VPAP™ III ST-A	N/V	X	MIRAGE ²	MIRAGE	MIRAGE	MIRAGE	MIRAGE	MIRAGE
VPAP™ IV / VPAP™ IV ST	N/V	Х	SWIFT ²	SWIFT	SWIFT	SWIFT	SWIFT	SWIFT
VPAP™ Adapt SV	N/V	X	N/V	N/V	N/V	N/V	N/V	N/V
VPAP™ Adapt (S9)	Pillows	X	Pillows ²	Pillows	Pillows	Pillows	Pillows	Pillows
VPAP™ Adapt (S9 PaceWave)	Pillows	X	Pillows ²	Pillows	Pillows	Pillows	Pillows	Pillows
VPAP™ Malibu	N/V	X	SWIFT ²	SWIFT	SWIFT	SWIFT	SWIFT	SWIFT
VPAP™ Auto 25	N/V	X	SWIFT ²	SWIFT	SWIFT	SWIFT	SWIFT	SWIFT
Auto 25 (S9)	Pillows	X	Pillows ²	Pillows	Pillows	Pillows	Pillows	Pillows
VPAP™ S / VPAP™ ST	N/V	X	SWIFT ²	SWIFT	SWIFT	SWIFT	SWIFT	SWIFT
VPAP™ S (S9) / VPAP™ ST (S9)	Pillows	X	Pillows ²	Pillows	Pillows	Pillows	Pillows	Pillows

Device	AirFit [™] P30i ¹	AirFit™ P10 For AirMini™	AirFit [™] P10 / AirFit [™] P10 for Her	Mirage Swift [™]	Mirage Swift™ II	Swift™ FX / Swift™ FX for Her / Swift™ FX Bella / Swift™ FX Bella Gray	Swift [™] LT	Swift™ LT for Her / Swift™ LT-F
VPAP™ ST-A (S9)	N/V	Х	Pillows ²	Pillows	Pillows	Pillows	Pillows	Pillows
VPAP™ ST (S9 iVAPS) / VPAP™ ST-A (S9 iVAPS)	N/V	Х	Pillows ²	Pillows	Pillows	Pillows	Pillows	Pillows
S9 VPAP™ Tx	Pillows ³	Χ	Pillows ²	N/V	Pillows	Pillows	Pillows	Pillows
VPAP™ Tx	N/V	Χ	PILLOWS ²	PILLOWS	PILLOWS	PILLOWS	PILLOWS	PILLOWS
Stellar™	N/V	Χ	PILLOWS ^{1, 2}	PILLOWS ¹	PILLOWS ¹	PILLOWS ¹	PILLOWS ¹	PILLOWS ¹
VS Ultra [™] / VS Integra [™]	N/V	Χ	N/V	N/V	N/V	N/V	N/V	N/V
VS III™	N/V	Χ	N/V	N/V	N/V	N/V	N/V	N/V
Astral™	N/V	X	N/V	Pillows	Pillows	Pillows	Pillows	Pillows

¹ SmartStop may not operate as described when using this mask with some CPAP or bi-level devices.

N/V = The mask has not been validated for this device.

 \checkmark = Compatible X = Not Compatible





² Accuracy of tidal volume/minute ventilation may reduce using the AirFit P10 mask in bilevel modes, particularly at higher pressure supports.

³ Excluding PAC and iVAPS modes. Only compatible for pressure range 4 - 25 cm H2O.

⁴ This combination of mask and device has been determined safe for use based on the evaluation of accuracy of monitoring of pressure, volume, minute ventilation and respiratory rate, and its alarms if applicable. Small variations in performance might be observed, particularly at higher pressure or flow. As per expected clinical practice, the effectiveness of ventilation and alarms should be verified by the clinician.